

Sludge, treated biowaste and soil- Determination of total organic carbon (TOC) by dry combustion

Einführendes Element — Haupt-Element — Ergänzendes Element

Élément introductif — Élément central — Élément complémentaire

ICS:

Descriptors:

Foreword

This draft standard has been prepared in the framework of project Horizontal.

This draft standard is a working document.

This standard is applicable and validated for several types of matrices. The table below indicates which ones.

Material	Validated for (type of sample, e.g. municipal sludge, compost)	Reference:
Sludge	Validated	EN 13137:2001
Soil	(Not yet validated)	(ISO 10694)
Biowaste	Not yet validated	
Sediment	Validated	EN 13137:2001
Waste	Validated	EN 13137:2001

Contents

Page

1	Scope.....	6
2	Normative references	6
3	Terms and definitions.....	6
3.1	Total carbon (TC).....	6
3.2	Total inorganic carbon (TIC).....	6
3.3	Total organic carbon (TOC).....	6
4	Safety remarks	7
5	Principle	7
5.1	Method A (indirect procedure).....	7
5.2	Method B (direct procedure)	7
5.3	Applicability of Methods A or B.....	7
6	Interferences and sources of errors	8
7	Reagents	8
7.1	General	8
7.2	Calcium carbonate	8
7.3	Sodium carbonate	8
7.4	Tetrasodium ethylenediamine tetraacetate-tetra-hydrate,	8
7.5	Potassium hydrogen phthalate.....	8
7.6	Acetanilide,.....	8
7.7	Atropine.....	9
7.8	Spectrographic graphite powder.....	9
7.9	Sodium salicylate	9
7.10	Aluminium oxide.....	9
7.11	Control mixture A	9
7.12	Control mixture B	9
7.13	Non-oxidizing mineral acid.....	9
7.14	Synthetic air, nitrogen, oxygen or argon	9
8	Apparatus	9
8.1	Homogenisation device,.....	9
8.2	Analytical balance	9
8.3	Equipment for determination of carbon in solids.....	9
8.4	Purging unit for TIC determination	10
8.5	Boats or crucibles	10
9	Sampling and sample pre-treatment.....	10
9.1	Sampling.....	10
9.2	Sample pre-treatment (to be uniformly coordinated with hor standard on pre-treatment).....	10
10	Procedure - Method A (Indirect method).....	10
10.1	Determination.....	10
10.1.1	General	10
10.1.2	Determination of the TC	11
10.1.3	Determination of the TIC	11
10.2	Calibration	11
10.3	Control measurements.....	12
10.4	Calculation and expression of results.....	12
11	Procedure Method B (direct method)	14
11.1	Determination.....	14
11.1.1	General	14

11.1.2	Removal of the inorganic carbon and determination of TOC	14
11.2	Calibration	15
11.3	Control measurements.....	15
11.4	Calculation and expression of results.....	15
12	Performance characteristics.....	16
13	Test report	17
Annex A	(informative) Validation of methods	18
A.1	Influence of temperature and modifiers on the decomposition of barium carbonate as an example for a refractory compound	19
A.2	Influence of aluminium oxide or sodium sulphate used for sample preparation on the recovery of TOC	19
A.3	Influence of TIC/TOC ratio on the recovery and the coefficient of variation.....	21
A.4	Method A: recovery of TOC for the control mixture A (7.11).....	23
A.5	Method B: influence of the temperature during the removal of inorganic carbon on the recovery of TOC	23
Annex B	(informative) The modular horizontal system.....	37
Annex C	(informative) Information on project Horizontal	38

Introduction

This draft standard is developed in the project 'Horizontal'. It is the result of a desk study of standards for determination of total organic carbon. The desk study aimed at evaluating the existing standards on determination of total organic carbon in sludge, soil, treated biowaste and neighbouring fields. After discussion with all parties concerned in CEN the standard has been developed further as a modular horizontal method and has been evaluated within the project 'Horizontal'

The texts of the Chapters 1 to 13 are normative; annexes are normative or informative, as stated in the top lines of the annexes.

1 Scope

This European Standard specifies two methods for the determination of total organic carbon (TOC) in sludge, sediment, biowaste and soil samples containing more than 1 g carbon per kg of dry matter (0,1%).

Coal and charcoal (elemental carbon) and inorganic carbon compounds except of carbonates will be determined as organic carbon when present in the sample.

(This standard based on EN 13137:2001 may include waste samples (and is prepared for) – depending on appropriate decision in TC 292 WG 5)

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text, and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

ISO 8466-1, Water quality - Calibration and evaluation of analytical methods and estimation of performance characteristics - Part 1: Statistical evaluation of the linear calibration function.

EN XXXXX, Soil, sludge, sediment and bio waste (and waste) – Physical characteristics - Determination of dry matter – Gravimetric method - Gravimetric method.

EN XXXXX, Soil, sludge, sediment and bio waste (and waste) – Physical characteristics - Determination of loss on ignition

EN XXXXX, Solid material - Sample pre-treatment - title subject to changes

EN XXXXX, Solid material – Sampling.... title subject to changes

EN XXXXX, Waste - Sample pre-treatment - title subject to changes

EN XXXXX, Waste – Sampling.... title subject to changes

ISO 10693, Soil quality – Determination of carbonate content – Volumetric method

ISO 3733, Petroleum products and bituminous materials - Determination of water - Distillation method.

ISO 6296, Petroleum products - Determination of water - Potentiometric Karl Fischer titration method.

3 Terms and definitions

For the purposes of this European Standard, the following terms and definitions apply:

3.1 Total carbon (TC)

the quantity of carbon present in the sample in the form of organic, inorganic and elemental carbon according to this standard

3.2 Total inorganic carbon (TIC)

the quantity of carbon that is liberated as carbon dioxide by acid treatment according to this standard

3.3 Total organic carbon (TOC)

the quantity carbon that is converted into carbon dioxide by combustion according to this standard and which is not liberated as carbon dioxide by acid treatment according to this standard

4 Safety remarks

Samples may be liable to fermentation and may be infectious. Due to this it is recommended to handle these samples with special care. The gases, which may occur due to the microorganisms activities, are potentially flammable. Excessive pressure build-up may cause the sample container to burst, potentially resulting in the formation of infectious aerosols and contaminated shrapnel.

Harmful compounds may arise during the combustion process and during the acid treatment. The user has to take appropriate precautions (e.g. activated carbon filters) to avoid these getting into the laboratory environment.

Samples with a high organic content may explode at introduction into the furnace. Using less sample material or covering the sample with inert material can reduce this risk.

5 Principle

The TOC can be measured either by Method A (indirect procedure) or by Method B (direct procedure).

5.1 Method A (indirect procedure)

In this procedure the TOC is obtained by the difference between the results of the measurements of TC and TIC.

The total carbon (TC) present in the un-dried sample or dried sample is converted to carbon dioxide by combustion in an oxygen-containing gas flow free of carbon dioxide. For soil dried samples are used. To ensure complete combustion, catalysts and/or modifiers can be used. The released amount of carbon dioxide is measured by infrared spectrometry, gravimetry, coulometry, conductometry, thermal conductivity detection, flame ionisation detection after reduction to methane, or other suitable techniques.

The TIC is determined separately from another sub-sample by means of acidification and purging of the released carbon dioxide. The carbon dioxide is measured by one of the techniques mentioned above. Alternatively, for soil the total organic carbon content may be calculated by determining the total carbon content and subtracting the carbon present as carbonate, which can be determined according to ISO 10693 (volumetric method).

5.2 Method B (direct procedure)

In this procedure the carbonates present in the un-dried or dried sample are previously removed by treating the sample with acid. The carbon dioxide released by the following combustion step is measured by one of the techniques mentioned in 5.1 and indicates the TOC directly.

5.3 Applicability of Methods A or B

Methods A and B have the same applicability in the terms of TOC content and/or TIC to TOC ratio. In samples with relatively high inorganic carbon contents method B is preferred.

Method B may lead to incorrect results in following cases:

- the sample contains volatile substances that evaporate during the acidification (e.g. volatile hydrocarbons from sludge of oil separators);
- side reactions between the sample and the acid take place (e.g. decarboxylation, volatile reaction products).

The quality of results of Method B is dependent on experience and practise, especially regarding the steps before the determination of TOC. Use of automatic dispensing units regarding removal of carbonates prior to determination of TOC may improve performance of Method B.

6 Interferences and sources of errors

Volatile organic substances may be lost during sample preparation. If necessary, the carbon content resulting from volatile organic substances shall be determined separately.

Depending on the laboratory experience on samples with high carbonate contents the procedures may lead to unreliable TOC results if the TIC to TOC ratio is very high (e.g. ≥ 10).

Depending on the detection method used, different interferences may occur, for instance:

- the presence of cyanide can interfere with the coulometric detection of TIC by modifying the pH value (dissolution of HCN);
- high content of halogenated compounds may lead to an overestimation of TOC when coulometric detection is used; in some cases the classical silver or copper trap can be insufficient to absorb all halides.

When present, elemental carbon, carbides, cyanides, cyanates, isocyanates, isothiocyanates and thiocyanates are determined as organic carbon using the methods described in this standard. An interpretation of the measured value may therefore be problematical in cases where the sample contains relevant levels of the above-mentioned components. If needed, these components shall be determined separately by means of a suitable validated procedure and be recorded in the test report.

Elementary carbon, determined separately, may be subtracted if required for the sample. If this is done this shall be reported by the laboratory.

7 Reagents

7.1 General

All reagents used shall be at least of analytical grade and suitable for their specific purposes.

Hygroscopic substances shall be stored in a desiccator.

7.2 Calcium carbonate

Calcium carbonate, CaCO_3

7.3 Sodium carbonate

Sodium carbonate Na_2CO_3 , anhydrous

7.4 Tetrasodium ethylenediamine tetraacetate-tetra-hydrate,

$\text{Na}_4\text{-EDTA}\cdot 4\text{H}_2\text{O}$ - $\text{C}_{10}\text{H}_{12}\text{N}_2\text{O}_8\text{Na}_4\cdot 4\text{H}_2\text{O}$, heated at 80°C for two hours.

NOTE Other forms of $\text{Na}_4\text{-EDTA}$ hydrates may be used if the water content is exactly known. Then the composition of the control mixtures has to be recalculated accordingly (see also 7.11 and 7.12).

7.5 Potassium hydrogen phthalate

Potassium hydrogen phthalate, $\text{C}_8\text{H}_5\text{O}_4\text{K}$

7.6 Acetanilide,

Acetanilide, $\text{C}_8\text{H}_9\text{NO}$

7.7 Atropine

Atropine, $C_{17}H_{23}NO_3$

7.8 Spectrographic graphite powder

Spectrographic graphite powder, C

7.9 Sodium salicylate

Sodium salicylate, $C_7H_5O_3Na$

7.10 Aluminium oxide

Aluminium oxide, Al_2O_3 , neutral, granular size < 200 μm , annealed at 600°C

7.11 Control mixture A

Control mixture A prepared from sodium carbonate (7.2), $Na_4-EDTA \cdot 4H_2O$ (7.4) and aluminium oxide (7.10) in a mass ratio of 2,34: 1,00:1,97.

The mixture shall be homogenized. It should contain 50,00 g/kg TIC and 50,00 g/kg TOC (e.g. 44,13 g of sodium carbonate, 18,83 g of $Na_4-EDTA \cdot 4H_2O$, 37,04 g of aluminium oxide).

7.12 Control mixture B

Control mixture B prepared from sodium salicylate (7.9), calcium carbonate (7.2), $Na_4-EDTA \cdot 4H_2O$ (7.4) and aluminium oxide (7.10) in a mass ratio of 1,00:4,36:1,97:8,40.

The mixture shall be homogenized. It should contain 33,3 g/kg TIC and 66,6 g/kg TOC (e.g. 6,36 g of sodium salicylate, 27,78 g of calcium carbonate, 12,50 g of $Na_4-EDTA \cdot 4H_2O$, 53,36 g of aluminium oxide).

7.13 Non-oxidizing mineral acid

Non-oxidizing mineral acid used for carbon dioxide expulsion, e.g. phosphoric acid H_3PO_4 (w = 85%)

Note Due to possible corrosion by hydrochloric acid, phosphoric acid is preferred.

7.14 Synthetic air, nitrogen, oxygen or argon

Synthetic air, nitrogen, oxygen, or argon, free of carbon dioxide and organic impurities in accordance with the manufacturer's instructions

8 Apparatus

8.1 Homogenisation device,

-for example mixers, stirrers, grinders, mills

8.2 Analytical balance

-accurate to at least 0,5% of test portion weight

8.3 Equipment for determination of carbon in solids

- with accessories

8.4 Purging unit for TIC determination

- to Method A only

8.5 Boats or crucibles

-made of e.g. ceramics, silica glass, silver or platinum

Note: Tin and nickel crucibles are not acid-resistant.

9 Sampling and sample pre-treatment

9.1 Sampling

Sampling are carried out according hor standard ENxxxx and/or ENyyyy ...

The samples are collected in glass or other suitable containers. Biologically active samples should be analysed immediately or stored at maximum - 18°C. For biologically inactive samples special preservation may not be necessary.

9.2 Sample pre-treatment (to be uniformly coordinated with hor standard on pre-treatment)

Pre-treatment of the sample is carried out according to horizontal standard on pre-treatment ENzzzzz and/or ENyyyyy

The samples supplied for analysis should be as homogeneous as possible. For soils, dry samples shall be used. Depending on the nature and appearance of the sample different procedures can be used, for example:

- a) Solid samples may be directly comminuted (avoiding heat) and reduced to a granular powder, preferable particle size less than 200 µm.
- b) Moist or paste-like samples may be mixed with aluminium oxide (7.10) until granular material is obtained and then comminuted, preferable particle size less than 200 µm. In this case the ratio of aluminium oxide to sample shall be considered in the calculation of TOC (10.4 or 11.4).

If samples contain - according to the accuracy of the method - negligible amounts of volatile compounds except water, the samples may be dried at 105°C before homogenisation. For analysing liquid sludge, especially sewage sludge, freeze-drying may also be used. In case of soil air-drying may be applied. In these cases the test report has to contain a clause: "sample dried at 105°C", "sample dried by freeze drying" (13d) respectively "samples air-dried".

The determination of the water content (if needed) shall be performed on a separate sub sample.

For samples containing no or negligible amounts of volatile organic compounds the water content is calculated from the determination of the dry matter according to EN XXXXX. Solid material.....

In case of samples containing volatile organic compounds the water content should be determined in another way, for instance in accordance with ISO 3733 (distillation method) or ISO 6296 (Karl-Fischer method).

10 Procedure - Method A (Indirect method)

10.1 Determination

10.1.1 General

This European Standard gives no recommendation concerning the construction of the apparatus and method of operation.

The operational characteristics should be selected and checked in accordance with the manufacturer's instructions.

The weight of the test portion should be as large as possible and shall be chosen so that the liberated quantity of carbon dioxide lies within the working range of the equipment/calibration.

10.1.2 Determination of the TC

The sample prepared according to 9.2 is weighed into a suitable vessel (boat or crucible made of e.g. ceramics, silica glass or platinum). To minimize carbon blank values the vessel may be pre-treated by heating (in a muffle-oven or the TC apparatus itself).

The sample is burned or decomposed in a flow of carrier gas containing oxygen (7.14).

The combustion temperature has to be high enough to convert all carbon completely to carbon dioxide. For samples containing carbonates, which are difficult to decompose, e.g. barium carbonate, the release of the carbon dioxide may be improved by increasing the temperature or by the use of modifiers, e.g. tin, copper.

The temperature range of commercially available instruments is between 900°C and 1500°C.

During the combustion of reactive samples, covering the sample with inert material, e.g. silica sand, may prevent detonation or fuming.

The carbon dioxide released during the analysis is measured using one of the detection methods in accordance with clause 4 and is expressed as carbon.

10.1.3 Determination of the TIC

The sample prepared according to 9.2 is weighed into the purging vessel (8.4).

The system is closed gas-tight and flushed with carrier gas until no more carbon dioxide from ambient air is present. Then acid (7.13) is added and the carbon dioxide is stripped by purging or stirring and/or heating. The released carbon dioxide is transferred to the detector by the carrier gas.

The addition of anti-foaming agents e.g. silicone oil may be helpful in the case of strongly foaming samples.

The addition of wetting agents e.g. surfactants may improve wetting of the surface of the sample.

The carbon dioxide released during the gas evolution is immediately measured using one of the detection procedures in accordance with clause 4 and is expressed as carbon.

TIC may alternatively be determined by using ISO 10693.

10.2 Calibration

If a relative method is used for detection, e.g. infrared detection, calibration is necessary.

Examples of calibration substances suitable for TC are calcium carbonate (7.2), potassium hydrogen phthalate (7.5), acetanilide (7.6), atropine (7.7), spectrographic graphite powder (7.8). Sodium carbonate (7.3) and Na₄-EDTA (7.4) as well as all compounds with EDTA-structure shall not be used for calibration as they are used as control substances.

Sodium carbonate (7.3) or calcium carbonate (7.2) are suitable for the calibration of TIC.

Other calibration substances may be used provided their suitability is checked.

The following procedure should be adopted during calibration:

Establish the preliminary working range.

- Measure a minimum of five (recommended ten) standard samples at least in triplicate.
- The concentration of these standard samples shall be distributed evenly over the working range.
- Calculate mean values for each concentration.
- Carry out a linear regression analysis with the mean values and test the linearity of the calibration function using ISO 8466-1.

The function shall be linear. Otherwise the working range must be restricted to the linear range.

If an absolute method is used for detection, e.g. coulometry, only control measurements according to 10.5 have to be carried out.

This calibration should be carried out for initial validation purposes or after major changes of the equipment.

10.3 Control measurements

Control measurements shall be performed to check that the equipment is functioning correctly. They should be carried out regularly using the control mixture A (7.11) for the procedures according to 10.1.2. (TC) and 10.1.3 (TIC). Analysis of one concentration from the middle of the respective working range, possible repeated two or three times, is sufficient. For the TC and TIC the mean recovery has to be between 90% and 110% with a coefficient of variation $\leq 5\%$.

The use of control charts for documentation of analytical quality is strongly recommended. When X-charts and r-charts are used, duplicate measurements of one control mixture sample per batch is sufficient.

A blank value shall be determined for all equipment and reagents used. It shall be taken into account if necessary.

If the required recoveries are not achieved, the following measures may be helpful:

TC analysis

- checking the homogeneity of the control mixture
- checking the calibration
- increasing the temperature during release of carbon dioxide
- using modifiers
- optimising the stirring speed and/or the gas flow in the purging vessel
- improving the gas exchange in the purging vessel
- avoiding condensation in the system.

10.4 Calculation and expression of results

The TC and TIC mass contents of the samples prepared according to clause 9 are calculated from

- calibration function and sample mass if relative detection methods are used
- specific constants and sample mass if absolute detection methods are used.

The calculation of TOC is achieved from the difference of the mean values of TC and TIC according to equation (1). In case of mixing the sample with aluminium oxide according to 9.2b) a dilution factor following equation (2) has to be considered:

$$\varpi_{TOC} = f(\varpi_{TC} - \varpi_{TIC}) \quad (1)$$

$$f = \frac{m_s + m_a}{m_s} \quad (2)$$

where:

- ϖ_{TOC} is the TOC content as carbon in the original sample in grams per kilogram (g/kg);
- ϖ_{TC} is the mean value of the TC content as carbon in the sample prepared according to 9.2 in grams per kilogram (g/kg);
- ϖ_{TIC} is the mean value of the TIC content as carbon in the sample prepared according to 9.2 in grams per kilogram (g/kg);
- f is the dilution factor resulting from the sample preparation of the original sample according to 9.2b);
- m_s is the mass of the original sample (to be mixed with aluminium oxide according to 9.2b);
- m_a is the mass of aluminium oxide according to 9.2b).

The TOC value resulting from equation (1) is calculated on dry matter by equation (3). For this purpose the water content determined separately according to 9.2 is used:

$$\varpi_{TOC_{dm}} = \varpi_{TOC} \times \frac{100}{100 - W} \quad (3)$$

where:

- $\varpi_{TOC_{dm}}$ is the TOC content as carbon, calculated on dry matter basis in grams per kilogram (g/kg);
- ϖ_{TOC} is the TOC content as carbon in the original sample in grams per kilogram (g/kg);
- W is the water content of the original sample according to 9.2 as mass fraction in percent (%).

The TOC content is reported as carbon on a dry matter basis. Following equation (3) results are obtained in g/kg. They can be converted into other units by using appropriate factors.

11 Procedure Method B (direct method)

11.1 Determination

11.1.1 General

This European Standard gives no recommendation concerning the construction of the apparatus and method of operation.

The operational characteristics should be selected and checked in accordance with the manufacturer's instructions.

The weight of the test portion should be as large as possible and shall be chosen so that the liberated quantity of carbon dioxide lies within the working range of the equipment/calibration.

11.1.2 Removal of the inorganic carbon and determination of TOC

The sample prepared according to 9.2 is weighed into a suitable vessel (boat or crucible made of e.g. ceramics, silica glass or platinum). The vessel may be prepared by thermal treatment (in a muffle oven or the combustion apparatus itself) to minimize carbon blank values.

To remove the inorganic carbon prior to the determination of the TOC the sample is carefully treated with a small volume of non-oxidizing mineral acid (7.13). Add the acid very slowly (dropwise) to avoid foaming and splashing of the sample. Add as little acid as possible but enough to soak the entire sample and to remove the inorganic carbon completely.

Note. An automatic dispenser system allowing small increments of acid to be added at a time is recommended. E.g. 12 increments of 50 µl non-oxidizing acid, diluted with water (1:10) per 10 – 25 mg of sample.

Allow at least 4 hours for the complete removal of the carbon dioxide. Stirring of the sample may reduce time needed for oxidation.

If moistening with the acid is difficult, the sample may be dampened beforehand with as little water as possible.

The moisture may be partly removed before combustion. The temperature during this sample treatment shall not exceed 40°C.

The sample is transferred to the combustion unit and heated carefully to remove the moisture. This may be realized by programming the temperature of the furnace or by inserting the sample slowly into the combustion unit. Then the sample is burnt in the carrier gas containing oxygen (7.14).

The combustion temperature has to be high enough to convert the organic carbon completely to carbon dioxide. The use of modifiers e.g. tin, copper may increase the recovery.

The temperature range of commercially available instruments lies between 900°C and 1500°C.

During the combustion of reactive samples detonation or fuming may be prevented by covering the sample with inert material e.g. silica sand after removal of the inorganic carbon.

The total carbon dioxide released during the combustion is measured using one of the detection methods in accordance with clause 4 and is expressed as carbon.

Corrosion of the combustion device may occur as a result of the acid remaining in the sample. Salt deposits may contaminate the system.

11.2 Calibration

The calibration for TOC has to be done in accordance with the calibration for TC (10.2). The selection of the calibration substances is analogous.

11.3 Control measurements

Control measurements should be carried out each working day using the control mixture B (7.12) for the procedure according to 11.1.2. Triple analysis of one point in the middle of the working range is sufficient. The mean recovery for the complete TOC procedure has to be between 90% and 110% with a coefficient of variation $\leq 10\%$.

The use of control charts for documentation of analytical quality is strongly recommended. When X-charts and r-charts are used, duplicate measurements of one control mixture sample per batch is sufficient.

A blank value has to be determined for all equipment and reagents used. It shall be taken into account if necessary.

If the required recovery is not achieved, the following measures may be helpful:

TOC analysis

- checking the homogeneity of the control mixture;
- checking the calibration;
- increasing the combustion temperature;
- reducing the flow of the carrier gas;
- encouraging a turbulent flow in the combustion tube;
- using modifiers:
- using post-oxidation of the combustion gases by catalysts.

Removal of carbonates

- decreasing the drying temperature of the acidified sample;
- decreasing the drying time of the acidified sample;
- omitting the drying step.

11.4 Calculation and expression of results

The TOC mass contents of the samples prepared according to 9.2 a) or b) are calculated from:

- calibration function and sample mass if relative detection methods are used,
- specific constants and sample mass if absolute detection methods are used.

The TOC is calculated on dry matter basis by equation (4). For this purpose the water content determined separately according to 9.2 and, if necessary, the dilution factor resulting from sample preparation is used.

$$\varpi_{TOCdm} = f \times \varpi_{TOC} \times \frac{100}{100 - W} \quad (4)$$

where:

ϖ_{TOCdm}	is the TOC content as carbon, calculated on dry matter basis in grams per kilogram (g/kg);
f	is the dilution factor resulting from the sample preparation of the original sample according to 9.2b) and in accordance with equation (2);
ϖ_{TOC}	is the mean value of TOC content as carbon in the sample prepared according to 9.2 in grams per kilogram (g/kg);
W	is the water content of the original sample according to 9.2 as mass fraction in percent (%).

The TOC content is reported as carbon on a dry matter basis. Following equation (4) results are obtained in g/kg. They can be converted into other units by using appropriate factors.

12 Performance characteristics

For precision data see Annex A.

13 Test report

The test report shall contain at least the following details:

- a) reference to this European Standard and the Method used (A or B);
- b) all necessary information on the full identification of the sample;
- c) the result according to 10.4 or 11.4 respectively;
- d) information of determined and subtracted amount of elementary carbon, if relevant;
- e) details of all procedural steps which deviate from this standard together with all circumstances that may have influenced the result.

Annex A (informative)

Validation of methods

The performance data of method A and B given in Table 1 have been established in an European interlaboratory study on five samples of waste and sludge and one synthetic mixture carried out in 1999.

Table 1 - Performance data (method A and B)

Sample	p	N	O (%)	m (% w_{dr})	s_R (% w_{dr})	S_R (%)	s_r (% w_{dr})	S_r (%)
METHOD A								
S1 synthetic mixture ¹⁾	20	76	0	8,662	0,7614	8,79	0,1854	2,14
S2 filter cake	20	76	1	7,798	1,4925	19,14	0,1716	2,20
S3 bottom ash	19	73	5	3,631	0,5548	15,28	0,1205	3,32
S4 electro-plating sludge	18	71	8	3,333	0,4029	12,09	0,1303	3,91
S5 dredged sludge	19	71	0	7,559	1,8239	24,13	0,3409	4,51
S6 rubble	19	71	5	6,759	0,8408	12,44	0,2156	3,19
METHOD B								
S1 synthetic mixture ¹⁾	10	40	20	8,332	0,4874	5,85	0,2316	2,78
S2 filter cake	12	46	8	8,498	0,8354	9,83	0,2507	2,95
S3 bottom ash	11	42	16	3,779	0,9905	26,21	0,1629	4,31
S4 electro-plating sludge	10	40	20	3,413	0,4792	14,04	0,1782	5,22
S5 dredged sludge	13	50	0	7,903	1,6588	20,99	0,4939	6,25
S6 rubble	11	43	12	6,327	1,6482	26,05	0,2461	3,89
¹⁾ = 9,1% (theoretical value) p Number of laboratories N Number of observed values O Percentage of outliers m General mean w_{dr} Dry matter s_R Estimate of the reproducibility standard deviation s_r Estimate of the repeatability standard deviation S_R Estimate of the relative reproducibility standard deviation S_r Estimate of the relative repeatability standard deviation								

Additional results of inter-laboratory tests

A.1 Influence of temperature and modifiers on the decomposition of barium carbonate as an example for a refractory compound

Table A.1 -Influence of temperature and modifiers on the decomposition of barium carbonate

-	-	900°C	1100°C	1300°C
modifier	ratio BaCO ₃ :modifier	TC%		
no modifier	-	< 0,1	< 0,1	5,8
vanadium pentoxide	1:2	6,3	6,3	6,3
tungsten oxide	1:2	< 0,1	6,0	6,3
copper/tungsten	1:1:3	-	5,5	6,2
tin	1:10	6,0	-	6,0

Result:

The use of modifiers may increase the recovery of TC for carbonates that originally decompose at higher temperatures, see Table A.1.

A.2 Influence of aluminium oxide or sodium sulphate used for sample preparation on the recovery of TOC

A real, moist sample (filter cake) was examined by several laboratories. The laboratories used aluminium oxide as well as sodium sulphate for sample preparation (see 9.2). To get a granular powder, it was necessary to mix the moist or pasty sample with aluminium oxide in a ratio 1:3 or with sodium sulphate in a ratio 1:4 to 1:14, see Table A.2.

Table A.2 - Influence of aluminium oxide or sodium sulphate used for sample preparation on the recovery of TOC

Laboratory identification	TC %			TOC					
	original sample	sample+ Al ₂ O ₃	sample+ Na ₂ SO ₄	method A %			method B %		
original sample				sample+ Al ₂ O ₃	sample+ Na ₂ SO ₄	original sample	sample+ Al ₂ O ₃	sample+ Na ₂ SO ₄	
-									
1	11,7	12,8	11,3	-	-	-	11,4	12,3	11,0
2	11,8	11,9	11,0	11,5	11,6	10,9	-	-	-
3	11,6	11,9	-	11,4	11,6	-	11,7	11,7	-
4	12,0	12,1	11,6	11,8	11,8	11,5	12,0	11,3	-
5	11,1	11,6	11,4	10,8	11,5	11,4	-	-	-
6	11,9	12,0	10,6	11,6	11,8	10,5	-	-	-
7	12,9	13,8	10,7	-	-	-	9,7	12,7	5,8
8	11,8	11,7	10,7	11,7	11,4	10,6	-	-	-
9	11,6	11,1	10,6	11,5	11,0	10,6	-	-	-
mean	11,8	12,1	11,0	11,5	11,5	10,9	11,2	12,0	8,4
standard deviation	0,48	0,78	0,40	0,33	0,28	0,44	1,03	0,62	-
coefficient of variation	4%	6%	4%	3%	2%	4%	9%	5%	-

Results:

The use of sodium sulphate for sample preparation leads to

- high consumption of sodium sulphate necessary to get a granular powder
- a higher limit of determination
- problems in the use of quartz combustion tubes
- generally reduced results compared with those from the original untreated sample.

The use of aluminium oxide for sample preparation leads to similar results compared with those from the original untreated samples and does not show the disadvantages cited above.

A.3 Influence of TIC/TOC ratio on the recovery and the coefficient of variation

Synthetic mixtures containing different ratios of calcium carbonate, glucose and aluminium oxide (TIC to TOC ratios 1:1 to 50:1) were examined by several laboratories, see Table A.3.1 and A.3.2.

Table A.3.1 - Method A: influence of TIC/TOC ratio on the recovery and the coefficient of variation

	ratio TIC:TOC	Laboratory identification	1	2	3	4	5	6			
		expected	found						mean	recovery	coefficient of variation
TC measurements %	5:5	10	9,9	10,0	9,7	9,9	9,7	9,94	9,85	98%	1%
	5:1	6	6,1	6,0	5,9	5,9	5,9	5,92	5,94	99%	1%
	5:0,5	5,5	5,4	5,5	5,5	5,5	5,5	5,43	5,45	99%	0%
	5:0,1	5,1	5,1	5,1	5,0	5,0	5,0	4,99	5,02	98%	1%
TIC measurements %	5:5	5	4,4	5,0	4,8	4,9	4,8	4,89	4,78	96%	4%
	5:1	5	4,8	5,0	4,9	4,9	4,9	4,88	4,90	98%	1%
	5:0,5	5	4,9	5,0	4,8	5,1	5,0	4,86	4,93	99%	2%
	5:0,1	5	4,9	5,0	4,8	4,9	4,9	4,90	4,89	98%	1%
TOC results %	5:5	5	5,5	5,0	4,9	5,0	4,9	5,05	5,06	101 %	5%
	5:1	1	1,2	1,0	1,0	1,0	1,0	1,04	1,05	105%	9%
	5:0,5	0,5	0,6	0,5	0,7	0,4	0,4	0,57	0,52	104%	20%
	5:0,1	0,1	0,1	0,1	0,2	0,0	0,2	0,08	0,11	112%	51%

Table A.3.2 - Method B: influence of TIC/TOC ratio on the recovery and the coefficient of variation

	ratio TIC:TOC	laboratory identification	A	B	C	D	E	mean	recovery	coefficient of variation
		expected	found							
TOC measurements %	5:5	5	5,3	5,1	5,0	4,5	4,5	4,91	98%	7%
	5:1	1	1,4	1,1	1,0	0,9	0,9	1,05	105%	20%
	5:0,5	0,5	0,6	0,5	0,5	0,9	0,4	0,57	114%	32%
	5:0,1	0,1	0,0	0,1	0,1	0,3	0,1	0,14	137%	65%

Result:

Up to a TIC to TOC ratio of 10:1 good recoveries and coefficients of variation were obtained for both methods (A and B). Within this limit there is no significant difference between the two methods.

A.4 Method A: recovery of TOC for the control mixture A (7.11)

A synthetic mixture of sodium carbonate, Na₄EDTA-4H₂O and aluminium oxide (TOC = TIC = 5%) was analysed by seven laboratories using method A. Results, see Table A.4.

Table A.4 - Method A: recovery of TOC for the control mixture

Laboratory identification	TC %	Recovery %	TIC %	Recovery %	TOC %	Recovery %
1	9,4	94	4,7	94	4,8	95
2	9,9	99	4,8	95	5,2	104
3	9,4	94	5,0	99	4,4	88
4	9,8	98	4,9	98	4,9	98
5	10,1	101	4,9	97	5,3	106
6	9,8	98	4,9	97	5,0	100
7	9,6	96	4,9	99	4,6	92
expected	10	-	5	-	5	-
mean	9,71	97	4,84	97	4,87	97
standard deviation	0,28		0,10		0,31	
coefficient of variation	3%	-	2%	-	6%	-

Result: For method A the required recovery was generally achieved.

A.5 Method B: influence of the temperature during the removal of inorganic carbon on the recovery of TOC

The control mixture B (7.12) with an expected value of 6,66% was analysed by one laboratory using the procedure of method B (11.1.2).

Table A.5 - Method B: influence of temperature during the removal of inorganic carbon on the recovery of TOC

Temperature °C	TOC %	Recovery %
20	6,6	99
30	6,5	98
40	6,3	94
50	5,3	80
70	4,1	62

Result:

Temperatures higher than 40°C during the removal of inorganic carbon lead to poor recoveries of TOC, see Table A.5.

ANNEX B

Determination of total organic carbon (TOC) in solid samples using the suspension method

(This method is subject to validation by interlaboratory studies after which the status (as an integrated part of the standard) will be determined)

TOC determinations in solid samples such as soils and sediments can alternatively be carried out in suspensions. These suspensions are, in terms of analytical sample preparation, comparable to wastewater samples containing particulate matter. TOC determination is carried out after quantitative oxidation to CO₂ according to EN 1484.

1. Application range

This appendix describes the determination of TOC in solid samples in the range of 0.1 – 20 %. The measuring range can, in principle, be extended downwards or upwards. This is conditional upon calibration of the corresponding measuring ranges, taking into consideration the sensitivity of the analytical system, sample dilution as well as blank values. The following matrices were successfully analysed using the suspension method:

- a) Soils (sandy silt, topsoil, clay, shale)
- b) Sediments (marine sediments, river sediments)
- c) Suspended sediments
- d) Raw and secondary materials in cement production (limestone, raw meal, fine dust, paper fibres, oil shale, dry sludges, cement powder).

2. Basic principle of the method

The TOC suspension method is a special sample preparation procedure for TOC determination in solid samples. The sample material is ground to a fine powder using an appropriate method (for example ball mill), with particle size $\leq 130 \mu\text{m}$ which is subsequently suspended in diluted hydrochloric acid. It is very important to minimize sedimentation of the suspension. Particles are effectively reduced and suspended using a homogeniser suitable for small particle sizes. The suspensions must be stirred immediately prior to sampling by the analytical system. The analytical system must be suitable with respect to sample preparation and dilution of particulate matter according to EN 1484.

The determination of organic carbon in the suspensions is carried out using the NPOC method (Non Purgeable Organic Carbon). The samples are suspended directly in hydrochloric acid in order to convert the inorganic carbon compounds (TIC) into CO₂. Using an inert gas, the CO₂ is subsequently purged from the sample. After drying, the sample no longer contains volatile compounds. Therefore the following simplification can be used: NPOC = TOC.

3.3 Hydrochloric acid

The suspensions are prepared in a 0.22 N hydrochloric acid solution. For the preparation of this solution, hydrochloric p.a. with a concentration of up to 37 % was used. Ultrapure water was used for the dilutions.

3.4 Gases

Gases according to EN 1484

Auxiliary gases such as carrier gas or sparging gas must be virtually free from CO, CO₂ or hydrocarbons.

4. Apparatus

Laboratory apparatus according to EN 1484, 300 mL Erlenmeyer flask especially suited for sample preparation.

4.1 Homogenisation

A high-velocity homogeniser is required which includes a precision tool capable of reducing particles to sizes $\leq 15 \mu\text{m}$, as well as a magnetic stirrer unit. The latter can also be part of the analytical apparatus.

5. Implementation

5.1 Calibration

In principle, the instrument is calibrated according to the directions of the manufacturer. A calibration curve is obtained, where the TC (Total Carbon) is established using potassium hydrogen phthalate standard solution of suitable concentration. In order to be applicable to a wide concentration range, several calibration curves may be necessary. A calibration curve corresponds to a dilution series of at least 5 concentration levels. A potassium hydrogen phthalate standard solution is diluted with appropriate volumes of ultrapure water. When available, the automatic dilution function of the analytical apparatus can be used to prepare the dilutions. For example: The standard solution contains 100 mg C/L. The dilution factors are 10, 5, 3, 2 and 1. This results in a dilution series with concentrations of 10.0, 20.0, 33.3, 50.0 and 100 mg C/L. A calibration line is established by plotting the TC mass concentrations, in milligrams per litre carbon, against the instrument-specific measuring value (I). The slope of this line corresponds to the instrument response factor and its reciprocal value is the calibration factor (f).

5.1 Control experiments

Control experiments with respect to calibration are to be carried out according to EN 1484. In addition the suitability of the instrument with regard to the suspension method should be thoroughly tested using reference materials. Suitable reference materials are:

- Certified reference material NIST 1941b (sediment), NIST, USA
- PT sample QTM068MS (sediment), QUASIMEME, The Netherlands

5.2 Determination

200 mg (± 10 mg) of the dried and ground sample are weighed into a 300 mL Erlenmeyer flask, to which 200 mL of a 0.22 N hydrochloric acid solution (3.3) is subsequently added. If the sample consists mainly of carbonates (limestone), the amount of sample can be increased to a maximum of 2000 mg. In both cases, the volume percentage of the solid is considered to be negligible. The volume of the suspension is therefore 200 mL. This entire sample volume is subsequently homogenised over 3 min using a high-velocity homogeniser (precision tool!) at 17,000 – 18,000 rpm. A suspension is formed. After purging with an inert gas, the NPOC and TOC in the suspension is determined as mean value from a minimum of 4 single injections and subsequently calculated with reference to the original solid sample. The suspension is stirred during purging and immediately prior to sampling.

6. Evaluation of the results

The mass percentage of organic carbon in the solid sample is calculated as follows:

$$\text{TOCs} = \frac{V_{\text{sus}} * I * 100}{E_s * f} [\% \text{m/m}]$$

TOC_S: organic carbon in the solid sample [% m/m]

I: instrument-specific measuring value [#]

V_{SUS}: volume of the suspension [L]

E_S: weight of the solid sample [mg]

f: calibration factor [# L / mg].

6.1 Blank value

The TC blank values of the ultrapure water and the 0.22 N hydrochloric acid may not exceed 0.3 mg/L and 0.6 mg/L respectively. With respect to the solid samples, the blank value is therefore < 0.06 % [m/m]. The calculation must take into account that ultra pure water is also used in the calibration. Therefore the following applies:

$$b = \frac{(b_{\text{HCL}} - b_{\text{H}_2\text{O}}) * V_{\text{SUS}} * 100}{E_s} [\%m/m]$$

b: blank value with respect to the solid sample [% m/m]

b_{HCL}: blank value of the 0.22 N hydrochloric acid [mg/L]

b_{H2O}: blank value of the ultrapure water.

The determination of the blank values b_{H2O} and b_{HCL} requires an analytical instrument with sufficient detection sensitivity for this concentration range (determination limit ≤ 200 µg/L).

7. Method characteristics

Table 7.1: Measurement of a certified reference material

Reference material	NIST 1941b
Target value (mass %)	2.99 ± 0.24
Number of in-house measurements	50 (10 series of 5 injections)
Overall mean value (mass %)	2.96
Number of outliers (Grubbs)	0
Mean recovery (%)	99.0
Relative standard deviation	
Within series (%)	5.6
Between series (%)	4.1
Overall standard deviation (%)	4.6
Extended measuring uncertainty (k = 2)	24.1%

Validation document for the TOC suspension method

(according to document A0-2 in the German Standard Methods for the Examination of Water, Wastewater and Sludge)

1. General information on development of the method

1.1 Start and end date of the process

08/2004 - 02/2006

1.2 Chairman and vice chairman

1.3 Membership List

2. Application area

2.1 Parameters determined

TOC

2.2 Area of operations

2.2.1 Tested matrices

- a) Soils (sandy silt, topsoil, clay, shale)
- b) Sediments (marine sediments, river sediments)
- c) Suspended matter
- d) Raw and secondary materials in cement production (limestone, raw meal, fine dust, paper, fibres, oil shale, dry sludge, cement powder)

2.2.2 Tested and calibrated concentration range

0.1 - 20 % (m/m)

2.2.3 Further applications of the method

In principle, this method can be applied to TC and TN_b in solids.

3. Basic principles of the method

The TOC suspension method is a special sample preparation method for TOC determination in solid samples. The sample material is ground using an appropriate grinding method (for example ball mill) to a fine powder with particle size $\leq 130 \mu\text{m}$, which is subsequently suspended in dilute hydrochloric acid.

Minimising of sedimentation of the suspension is very important. Sediment particles are effectively suspended using a homogeniser suitable for small particle sizes. The suspensions must be stirred immediately prior to sampling by the analytical system. The analytical system must be suitable with respect to samples containing particulate matter according to EN 1484.

Determination of organic carbon in suspensions is carried out using the NPOC method (Non Purgeable Organic Carbon). The samples are suspended directly in hydrochloric acid in order to convert the inorganic carbon compounds (TIC) into CO₂. Using an inert gas, the CO₂ is subsequently purged from the sample. After drying, the sample no longer contains volatile compounds. The following simplification can therefore be used: NPOC = TOC.

4. Interferences

- Insufficient grinding, i.e. to attain particle sizes $< 130 \mu\text{m}$, can potentially lead to poor reproducibility
- Inclusions of, for instance, vitreous elemental carbon as well as carbonates. Fly ashes show significant differences in comparison with measuring methods for solids and suspensions.

5. Reagents, test organisms, instruments

5.1 Blank values

The ultra pure water used for dilution and the hydrochloric acid (p.a.) generate blank values. With respect to the solid sample, the blank value lies between 0.02 and 0.06 % carbon (m/m), provided that the instruments used are clean.

5.2 Requirements with respect to purity of reagents

32 % hydrochloric acid p.a., ultra pure water (carbon content $< 300 \mu\text{g/L}$), potassium hydrogen phthalate p.a.

5.3 Availability of reagents, standards and reference materials

- Certified reference material NIST 1941b (marine sediment), NIST, Gaithersburg, MD 20899-1070, USA
- Reference material QTM068MS (marine sediment), QUASIMEME, Wageningen, the Netherlands

5.4 Stability of reagents

See information and instructions in the EN Standards listed in section 15.

5.5 Test organisms

No test organisms were used.

5.6 Chromatographic separation phases

No separation phases were used.

5.7 Instruments

The detection sensitivity of the analytical instruments must be sufficiently high in order to measure the blank value. Determination limit: TC $< 200 \mu\text{g C/L}$.

5.8 Occupational safety and environmental protection/pollution control

See information and instructions in the DIN Standards listed in section 15.

6 Sampling and sample preparation

6.1 Sampling

See information and instructions in the EN standards listed in section 15.

6.2 Sample stability and sample preservation

See information and instructions in the EN Standards listed in section 15.

6.3 Occupational safety and environmental protection

See information and instructions in the EN Standards listed in section 15.

7. Implementation

7.1 Sample preparation enrichment conditions

200 mg (± 10 mg) of the dried and ground sample are weighed into a 300 mL Erlenmeyer, to which 200 mL of a 0.22 N hydrochloric acid solution is subsequently added. If the sample consists mainly of carbonates (limestone), the amount of sample can be increased to a maximum of 2000 mg. In both cases, the volume percentage of the solid is considered to be negligible. The volume of the suspension is therefore 200 mL. This entire sample volume is subsequently homogenised over 3 min using a high-velocity homogeniser (precision tool) at 17,000 – 18,000 rpm. A suspension is formed.

7.2 Sample measurement, measuring instrument parameters

After purging with an inert gas, the NPOC and TOC in the suspension is determined as mean value from 4 single injections and subsequently calculated with reference to the original solid sample. The suspension has been stirred during purging and immediately prior to withdrawing of an aliquot of the sample. Apart from this, the guidelines of the instrument manufacturer are standard.

8. Determination of the method characteristics

8.1 Variances in homogeneity

Data are available (F-test, PW = 1.17, KW = 18.51).

8.2 Linearity testing

F-test according to Mandel

Optimal regression model: linear (PW = 4.05 KW = 21.19).

8.3 Type of calibration


Linear regression

8.4 Calibration standards

Potassium hydrogen phthalate p.a. 100 mg/L as stock solution

8.5 Calibration data and functions

Number	x-value	f(x)-value
	mg/L	Area
1	2.0	15.763
2	5.0	36.180
3	10.0	72.250
4	20.0	146.300
5	33.3	246.950
6	50.0	373.400
7	100.0	734.150



Residues are normally distributed (R/s test, 99 %, PW = 2.89, KW = 2.26 – 3.33)

Residues do not show a trend (Neumann test, 99 %, PW = 2.03, KW = 0.61)

Residual standard deviation: 2.748

Standard deviation of the method: 0.373

$f(x)\text{-value} = 7.36 * x\text{-value} + 0.475$

R = 0.9999, N = 7

8.6 Detection, quantification and determination limit (according to DIN 32645)

Detection limit	1.1 mg/L	0.1 %
Quantification limit	2.2 mg/ L	0.2 %
Determination limit	3.9 mg/ L	0.4 %

Alternatively, calibrations were carried out in other concentration ranges with a possible determination limit of < 100 µg C/L, corresponding to < 0.1 % (calibration range up to 1000 µg C/L).

8.7 Recalibration

See information and instructions in the EN Standards listed in section 15.

9 Testing for trueness (accuracy of the mean)

9.1 Reference material

See information in section 5.

9.2 Standard addition method for the matrices

No standard addition method was deployed.

9.3 Recovery rates

Reference material NIST 1941b: 99.0 % for N = 50 measurements.

Reference material QTM068MS: 97.6 % for N = 30 measurements.

9.4 Comparison with results of other analytical procedures

Measurements of different types of soil

1 = Suspension, 2 = SSM

Mean 1	0.6%	0.7%	1.2%	1.7%
Mean 2	0.4%	0.8%	1.4%	2.0%
sd 1	0.001	0.000	0.001	0.001
sd 2	0.003	0.001	0.001	0.0004
N1	5	4	5	5
N2	5	5	5	5
t	1.233	1.364	2.304	2.004
t(f,P=95%)	2.306	2.365	2.306	2.306
t(f,P=99%)	3.355	3.499	3.355	3.355

10. Precision testing

10.1 Type of samples used

Reference materials, see section 5.

10.2 Statistical evaluation

Reference material 1	NIST 1941b
Target value (mass %)	2.99 ± 0.24
Number of in-house measurements	50 (10 series of 5 injections)
Time frame	April to August 2005
Overall mean value (mass %)	2.96
Number of outliers (Grubbs)	0
Mean recovery (%)	99.0
Relative standard deviations	
Within series (%)	5.6
Between series (%)	4.1
Overall standard deviation (%)	4.6

Reference material 2	QTM068MS
Target value (mass %)	2.23
Total error (%)	± 0.33
Number of measurements	30 (6 series of each 5 injections)
Time frame	May to June 2005
Overall mean value (mass %)	2.15
Number of outliers (Grubbs)	0
Mean recovery (%)	97.6
Relative standard deviations	
Within series (%)	5.3
Between series (%)	8.9
Overall standard deviation (%)	7.5

11. Robustness, stability

See section 4.

12. Characteristic method data from round robin tests

Three laboratories participated in the following tests on the TOC suspension method:

- a) ISE (International Soil Analytical Exchange, WEPAL, Wageningen University, Wageningen, NL)
- b) SETOC (International Sediment Exchange for Tests on Organic Contaminants, WEPAL, Wageningen University, Wageningen, NL)

12.1 Time frame of the round robin tests

- a) Time frame: October - December 2005; number of participating laboratories: 32
- b) Time frame: October - December 2005; number of participating laboratories: 17

12.2 Parameters analysed as described in section 2

TOC (mg/kg)

12.3 Reference materials used as described in section 5

- a) 100 g soil samples 1 - 4 (2 x sandy soil, 2 x clay)
- b) 100 g sediment samples 1 - 4 (3 x sediment, 1 x clay)

12.4 Matrices studied

See 12.3

12.5 Concentration ranges studied

Sample	TOC-values [mg/kg]			Z-Scores		
	Lab 1	Lab 2	Lab 3	Lab 1	Lab 2	Lab 3
SETOC-1	101	101	101.0	-0.56	-0.56	-0.56
SETOC-2	41.9	43.8	41.3	-0.31	0.16	-0.46
SETOC-3	44.0	46.5	42.4	0.38	2.04	-0.68
SETOC-4	57.9	56.6	56.9	-0.23	-0.82	-0.69
ISE-1	13.9	16.3	15.2	-0.43	1.78	0.77
ISE-2	18.3	17.8	19.6	0.52	0.03	1.78
ISE-3	56.8	54.5	56.1	-1.15	-2.93	-1.69
ISE-4	46.3	46.0	49.9	-2.84	-3.01	-0.80

Sample	Median	MAD	N	Mean	sR	rel sR
SETOC-1	104	3.00	13	103	4.40	4.2
SETOC-2	43.8	1.90	15	43.2	4.09	9.5
SETOC-3	43.7	0.83	12	43.4	1.51	3.5
SETOC-4	58.1	1.50	13	58.4	2.21	3.8
ISE-1	14.2	0.70	29	14.4	1.09	7.6
ISE-2	17.8	0.60	26	17.8	1.03	5.8
ISE-3	58.2	0.80	24	58.3	1.29	2.2
ISE-4	51.3	1.21	26	51.3	1.76	3.4

MAD = Median of absolute deviation

sR = Reproducibility

rel sR = Relative reproducibility

12.6 Outlier rate

No information from the PT provider

12.7 Relative repeatability

No information from the PT provider

12.8 Relative reproducibility

For relative reproducibility (rel sR) see 12.5

12.9 Comparison of results obtained by other analytical procedures

See section 9.4

13. Measurement uncertainty

13.1 Type of the investigation

Multiple measurements of certified reference materials in section 5.3

13.2 Typical result regarding specification of the matrix and concentration levels

$U(y) = 12,3 \% - 20,9 \%$

The measurement uncertainty indicated is an expanded uncertainty (coverage factor $k = 2$, confidence level 95 %) calculated for sediments in the concentration range 2 – 4 %.

14. Evaluation

14.1 Identification criteria

See information and instructions in the EN Standards listed in section 15.

14.2 Calculation and specification of the results

See information and instructions in the EN Standards listed in section 15.

15. Literature

EN 1484: 1997 (DEV H3) Guidelines for the determination of total organic carbon (TOC) and dissolved organic carbon (DOC).

EN 13137: 2001 Characterisation of waste – determination of total organic carbon (TOC) in waste, sludges and sediments.

Servos U, Gluschke M, Kramer T: Der TOC in Sedimenten und Böden – Fortschritte in der Analytik. Wasser, Luft und Boden (wlb), 10/2005, 46 – 49.

AnnexC
(informative)

The modular horizontal system

AnnexD
(informative)

Information on project Horizontal

Bibliography